

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4802

Name: RITTER 1  
Sterilizer ID: V2178198  
08/12/2024 09:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.1 F 0.0  
2:00 159.0 F 1.2  
4:00 174.4 F 3.3  
6:00 198.2 F 7.1  
8:00 211.9 F 11.3  
10:00 220.3 F 4.4  
12:00 229.1 F 8.9  
14:00 239.0 F 13.4  
16:00 239.5 F 10.3  
18:00 236.9 F 8.6  
20:00 248.3 F 14.2  
22:00 259.1 F 20.1  
24:00 268.8 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 271.8 F 28.6  
2:00 272.3 F 29.0  
3:00 272.5 F 29.0  
4:00 272.3 F 28.8

Min 271.1 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.1 F 0.0  
5:00 240.7 F 0.0  
10:00 219.7 F 0.0  
15:00 235.6 F 0.0  
20:00 194.2 F 0.0  
25:00 183.5 F 0.0  
30:00 179.3 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 25:28  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:04:28  
CYCLE COMPLETE

08/12/2024 10:56 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_