
Midmark M11 - v1.0.0
Total Cycles: 2913

Name: RITTER 1
Sterilizer ID: V2178198
02/16/2023 10:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.6 F 0.0
2:00 164.5 F 3.1
4:00 179.9 F 7.5
6:00 202.4 F 12.2
8:00 221.4 F 5.7
10:00 231.3 F 10.9
12:00 232.0 F 7.9
14:00 241.5 F 11.5
16:00 244.6 F 12.7
18:00 256.6 F 19.1
20:00 267.5 F 25.7
22:00 273.0 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 29.1
1:00 272.6 F 29.8
2:00 272.4 F 29.2
3:00 272.4 F 29.6
4:00 272.6 F 29.5

Min 271.8 F 29.0
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.4 F 0.0
5:00 231.6 F 0.0
10:00 235.2 F 0.0
15:00 223.9 F 0.0
20:00 216.5 F 0.0
25:00 208.8 F 0.0
30:00 213.0 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 22:11
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:00:29
CYCLE COMPLETE

02/16/2023 11:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____