
Midmark M11 - v1.0.5
Total Cycles: 2948

Name: RITTER2
Sterilizer ID: V2313228
09/29/2023 08:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.9 F 0.0
2:00 72.4 F 0.0
4:00 109.6 F 0.0
6:00 138.0 F 0.0
8:00 155.2 F 0.4
10:00 174.3 F 1.5
12:00 192.1 F 4.4
14:00 210.8 F 8.1
16:00 219.9 F 11.9
18:00 224.4 F 5.9
20:00 234.4 F 10.3
22:00 238.1 F 8.9
24:00 242.4 F 12.4
26:00 240.9 F 11.0
28:00 253.1 F 17.1
30:00 263.7 F 23.4
32:00 272.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.4 F 29.2
2:00 271.5 F 28.9
3:00 272.2 F 29.1
4:00 271.5 F 28.9

Min 271.4 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 207.5 F 0.0
10:00 211.1 F 0.0
15:00 207.9 F 0.0
20:00 208.1 F 0.0
25:00 209.2 F 0.0
30:00 208.2 F 0.0

Drying Complete

FILLING: 0:48
HEATING: 32:13
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:10:23
CYCLE COMPLETE

09/29/2023 10:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____