
Midmark M11 - v1.0.0
Total Cycles: 1635

Name: RITTER 1
Sterilizer ID: V2178198
02/15/2022 11:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.1 F 0.0
2:00 183.1 F 0.0
4:00 208.8 F 0.2
6:00 217.0 F 1.3
8:00 228.1 F 5.0
10:00 239.1 F 9.5
12:00 241.2 F 10.1
14:00 247.9 F 13.0
16:00 247.0 F 13.1
18:00 246.6 F 12.9
20:00 260.6 F 20.6
22:00 272.7 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.6
1:00 272.1 F 27.7
2:00 272.8 F 28.2
3:00 272.1 F 28.1
4:00 272.1 F 27.8

Min 271.6 F 27.5
Max 272.9 F 28.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 237.2 F 0.0
10:00 226.6 F 0.0
15:00 225.7 F 0.0
20:00 222.4 F 0.0
25:00 233.4 F 0.0
30:00 231.3 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:21
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:00:26
CYCLE COMPLETE

02/15/2022 12:39 PM
ACCEPTED: X
REJECTED:
Operator: U02