
Midmark M11 - v1.0.0
Total Cycles: 2019

Name: RITTER 1
Sterilizer ID: V2178198
06/09/2022 02:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.5 F 0.0
2:00 193.8 F 4.5
4:00 219.3 F 9.7
6:00 226.0 F 5.8
8:00 236.6 F 11.3
10:00 234.8 F 8.5
12:00 239.8 F 9.8
14:00 247.2 F 13.8
16:00 259.7 F 20.7
18:00 270.9 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.2 F 29.4
2:00 271.7 F 29.0
3:00 272.0 F 28.8
4:00 272.7 F 29.2

Min 270.9 F 28.1
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 220.7 F 0.0
10:00 245.4 F 0.0
15:00 236.1 F 0.0
20:00 224.6 F 0.0
25:00 223.4 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 2:27
HEATING: 18:36
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:15
CYCLE COMPLETE

06/09/2022 03:30 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____