
Midmark M11 - v1.0.5
Total Cycles: 1837

Name: RITTER2
Sterilizer ID: V2313228
12/14/2022 10:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.4 F 0.0
2:00 170.4 F 0.0
4:00 180.3 F 0.0
6:00 194.6 F 4.3
8:00 214.4 F 9.3
10:00 230.4 F 9.3
12:00 236.7 F 10.6
14:00 234.2 F 8.3
16:00 240.6 F 10.4
18:00 248.4 F 14.5
20:00 260.9 F 21.5
22:00 271.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.5 F 29.3
2:00 271.7 F 28.5
3:00 272.2 F 29.4
4:00 272.3 F 29.1

Min 271.4 F 28.5
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.4 F 0.0
5:00 213.0 F 0.0
10:00 228.1 F 0.0
15:00 241.0 F 0.0
20:00 221.4 F 0.0
25:00 217.4 F 0.0
30:00 220.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 22:38
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:00:35
CYCLE COMPLETE

12/14/2022 11:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____