
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002412
12:10 10/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 1:59
STERILIZING 2:59
136.4 C 222kPa 2:59
136.3 C 224kPa 3:29
136.3 C 223kPa 3:59
136.4 C 224kPa 4:29
136.2 C 224kPa 4:59
136.3 C 224kPa 5:29
136.2 C 223kPa 5:59
136.1 C 223kPa 6:29
136.1 C 223kPa 6:29
VENTING 6:29
AIR DRYING 7:18
CYCLE COMPLETE 15:15

Digital Signature #
C492CCB4901EC8C7
