
Midmark M11 - v1.0.5
Total Cycles: 163

Name: RITTER2
Sterilizer ID: V2313228
07/05/2021 02:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.8 F 0.0
2:00 165.9 F 3.0
4:00 184.5 F 7.5
6:00 206.8 F 12.7
8:00 222.8 F 7.7
10:00 235.8 F 13.1
12:00 238.6 F 11.1
14:00 238.4 F 10.2
16:00 251.9 F 16.8
18:00 263.5 F 23.7
20:00 271.9 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.7 F 29.7
2:00 271.5 F 29.0
3:00 272.6 F 29.9
4:00 272.1 F 29.3

Min 271.4 F 28.9
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.9 F 0.0
Dry Cycle Aborted

FILLING: 1:57
HEATING: 20:08
STERILIZING: 4:00
VENTING: 3:12
DRYING: 0:37
TOTAL CYCLE: 00:29:54

07/05/2021 03:10 PM
ACCEPTED: X
REJECTED:
Operator: U01