
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001812
8:32 08/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:14
PRESSURIZING 2:20
STERILIZING 3:21
136.0 C 221kPa 3:21
136.2 C 223kPa 3:51
136.4 C 222kPa 4:21
135.9 C 221kPa 4:51
136.3 C 225kPa 5:21
136.4 C 226kPa 5:51
136.1 C 223kPa 6:21
136.2 C 222kPa 6:51
136.2 C 222kPa 6:51
VENTING 6:51
AIR DRYING 7:43
CYCLE COMPLETE 7:58

Digital Signature #
C84BF1114F2D9E36
