
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001130
9:22 21/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:04
PRESSURIZING 4:05
STERILIZING 5:48
136.5 C 225kPa 5:48
136.2 C 223kPa 6:18
136.4 C 225kPa 6:48
136.2 C 223kPa 7:18
136.3 C 224kPa 7:48
136.1 C 223kPa 8:18
136.3 C 224kPa 8:48
136.4 C 224kPa 9:18
136.4 C 224kPa 9:18
VENTING 9:18
AIR DRYING 10:22
CYCLE COMPLETE 13:41

Digital Signature #
61AE1E377930F9D9
