

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1403

Name: RITTER 1  
Sterilizer ID: V2178198  
11/29/2021 04:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.1 F 0.0  
2:00 191.6 F 3.9  
4:00 203.7 F 8.5  
6:00 220.6 F 11.0  
8:00 227.7 F 8.5  
10:00 238.9 F 11.7  
12:00 241.3 F 11.4  
14:00 239.0 F 10.0  
16:00 253.2 F 17.0  
18:00 265.2 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.3 F 29.0  
2:00 272.5 F 29.0  
3:00 272.3 F 28.6  
4:00 271.5 F 28.2  
  
Min 271.1 F 28.0  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.1 F 0.0  
5:00 244.1 F 0.0  
10:00 224.5 F 0.0  
15:00 218.3 F 0.0  
20:00 197.0 F 0.0  
25:00 203.1 F 0.0  
30:00 189.7 F 0.0  
Drying Complete

FILLING: 2:19  
HEATING: 19:53  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:24  
CYCLE COMPLETE

11/30/2021 07:54 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02