
Midmark M11 - v1.0.5
Total Cycles: 445

Name: RITTER2
Sterilizer ID: V2313228
10/13/2021 11:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.2 F 0.0
2:00 173.2 F 3.2
4:00 182.4 F 7.7
6:00 204.2 F 13.1
8:00 224.7 F 8.4
10:00 236.1 F 11.4
12:00 240.0 F 11.9
14:00 239.9 F 10.8
16:00 253.2 F 17.6
18:00 264.8 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.2 F 29.7
2:00 271.2 F 29.1
3:00 272.4 F 30.4
4:00 272.4 F 29.8

Min 270.9 F 29.0
Max 273.6 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.7 F 0.0
5:00 214.9 F 0.0
10:00 224.2 F 0.0
15:00 178.9 F 0.0
20:00 170.1 F 0.0
25:00 156.8 F 0.0
Dry Cycle Aborted

FILLING: 1:40
HEATING: 19:49
STERILIZING: 4:00
VENTING: 3:09
DRYING: 28:53
TOTAL CYCLE: 00:57:31

10/13/2021 12:45 PM
ACCEPTED: X
REJECTED:
Operator: U01