

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2420

Name: RITTER2  
Sterilizer ID: V2313228  
05/26/2023 10:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 133.5 F 0.0  
2:00 134.5 F 0.0  
4:00 144.2 F 0.5  
6:00 172.1 F 4.1  
8:00 194.1 F 8.4  
10:00 212.2 F 13.0  
12:00 221.3 F 6.6  
14:00 232.9 F 11.1  
16:00 232.5 F 7.9  
18:00 244.8 F 12.9  
20:00 244.5 F 12.6  
22:00 256.3 F 18.9  
24:00 266.9 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.5 F 29.2  
2:00 271.5 F 28.6  
3:00 272.5 F 29.3  
4:00 271.7 F 28.6

Min 271.4 F 28.5  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.3 F 0.0  
5:00 208.7 F 0.0  
10:00 204.0 F 0.0  
15:00 203.5 F 0.0  
20:00 206.6 F 0.0  
25:00 210.2 F 0.0  
30:00 210.9 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 25:32  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:03:45  
CYCLE COMPLETE

05/26/2023 11:58 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_