
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001605
16:13 16/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:00
STERILIZING 2:59
136.0 C 221kPa 2:59
136.4 C 226kPa 3:29
136.4 C 226kPa 3:59
136.1 C 223kPa 4:29
136.4 C 225kPa 4:59
136.4 C 225kPa 5:29
136.3 C 223kPa 5:59
136.4 C 224kPa 6:29
136.4 C 224kPa 6:29
VENTING 6:29
AIR DRYING 7:22
CYCLE COMPLETE 7:27

Digital Signature #
3CD2CF2F2D34174E
