
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005639
12:12 31/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:11
STERILIZING 3:15
136.4 C 221kPa 3:15
136.4 C 221kPa 3:45
136.4 C 221kPa 4:15
136.4 C 223kPa 4:45
136.2 C 224kPa 5:15
136.0 C 221kPa 5:45
135.8 C 219kPa 6:15
136.4 C 224kPa 6:45
136.4 C 224kPa 6:45
VENTING 6:45
AIR DRYING 7:36
CASSETTE REMOVED
CYCLE COMPLETE 31:42

Digital Signature #
F55F75936839F6FF
