

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007638  
8:47 05/07/2024

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:20  
PRESSURIZING 3:23  
STERILIZING 4:33  
136.0 C 220kPa 4:33  
135.8 C 219kPa 5:03  
136.4 C 223kPa 5:33  
136.4 C 220kPa 6:03  
136.4 C 222kPa 6:33  
136.4 C 222kPa 7:03  
136.4 C 223kPa 7:33  
136.3 C 221kPa 8:03  
136.3 C 221kPa 8:03  
VENTING 8:03  
AIR DRYING 8:53  
CYCLE COMPLETE 14:13

-----

Digital Signature #  
6E66A1D219868DC5  
-----