
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002555
8:38 04/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:31
STERILIZING 3:40
136.0 C 220kPa 3:40
136.2 C 222kPa 4:10
136.4 C 225kPa 4:40
136.2 C 222kPa 5:10
136.4 C 224kPa 5:40
136.4 C 224kPa 6:10
136.4 C 220kPa 6:40
136.1 C 219kPa 7:10
136.1 C 219kPa 7:10
VENTING 7:10
AIR DRYING 8:03
CYCLE COMPLETE 8:46

Digital Signature #
E5413605F14035F7
