
Midmark M11 - v1.0.0
Total Cycles: 1048

Name: RITTER 1
Sterilizer ID: V2178198
07/22/2021 01:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.0 F 0.0
2:00 159.3 F 2.3
4:00 183.3 F 5.7
6:00 207.2 F 10.9
8:00 221.5 F 5.1
10:00 232.2 F 10.0
12:00 233.8 F 8.1
14:00 244.0 F 12.6
16:00 241.8 F 11.0
18:00 255.0 F 17.7
20:00 266.6 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.7 F 28.9
2:00 271.8 F 28.1
3:00 271.9 F 28.5
4:00 272.8 F 28.8

Min 271.2 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.6 F 0.0
5:00 233.5 F 0.0
10:00 201.7 F 0.0
15:00 201.4 F 0.0
20:00 173.4 F 0.0
25:00 153.1 F 0.0
30:00 145.2 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 21:35
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:03
CYCLE COMPLETE

07/22/2021 02:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____