
Midmark M11 - v1.0.0
Total Cycles: 4084

Name: RITTER 1
Sterilizer ID: V2178198
12/05/2023 11:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.8 F 0.0
2:00 152.1 F 2.0
4:00 161.9 F 5.4
6:00 186.6 F 9.3
8:00 202.9 F 10.4
10:00 217.5 F 6.2
12:00 228.3 F 10.5
14:00 233.2 F 7.7
16:00 241.7 F 12.2
18:00 240.4 F 10.6
20:00 253.1 F 17.0
22:00 264.4 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.0 F 28.9
2:00 272.6 F 28.9
3:00 271.7 F 28.4
4:00 271.7 F 28.7

Min 271.4 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.4 F 0.0
5:00 232.7 F 0.0
10:00 228.3 F 0.0
15:00 228.3 F 0.0
20:00 228.0 F 0.0
25:00 228.5 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:52
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:02:36
CYCLE COMPLETE

12/05/2023 12:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____