
Midmark M11 - v1.0.0
Total Cycles: 5233

Name: RITTER 1
Sterilizer ID: V2178198
01/10/2025 10:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.6 F 0.0
2:00 181.2 F 1.9
4:00 183.9 F 5.5
6:00 200.2 F 9.9
8:00 219.8 F 7.2
10:00 223.2 F 7.5
12:00 234.2 F 12.0
14:00 234.5 F 8.4
16:00 243.8 F 12.2
18:00 244.5 F 12.3
20:00 256.1 F 18.4
22:00 266.4 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 271.9 F 28.3
2:00 272.4 F 28.7
3:00 271.6 F 28.5
4:00 271.7 F 28.1

Min 271.2 F 28.0
Max 272.4 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.6 F 0.0
5:00 231.0 F 0.0
10:00 202.6 F 0.0
15:00 191.0 F 0.0
20:00 183.2 F 0.0
25:00 183.0 F 0.0
30:00 182.4 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 23:37
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:02:46
CYCLE COMPLETE

01/10/2025 11:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____