
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001882
13:08 23/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:49
STERILIZING 3:55
136.0 C 220kPa 3:55
136.2 C 223kPa 4:25
136.4 C 225kPa 4:55
136.4 C 223kPa 5:25
135.9 C 220kPa 5:55
136.1 C 222kPa 6:25
136.4 C 225kPa 6:55
136.4 C 225kPa 7:25
136.4 C 225kPa 7:25
VENTING 7:25
AIR DRYING 8:13
CYCLE COMPLETE 8:40

Digital Signature #
85360F5D339715C8
