
Midmark M11 - v1.0.5
Total Cycles: 2369

Name: RITTER2
Sterilizer ID: V2313228
05/15/2023 03:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.1 F 0.0
2:00 193.3 F 4.0
4:00 205.7 F 9.4
6:00 223.0 F 7.8
8:00 233.0 F 10.9
10:00 233.8 F 8.8
12:00 239.0 F 10.1
14:00 247.9 F 14.7
16:00 260.0 F 21.5
18:00 270.9 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.3 F 29.5
2:00 271.9 F 29.7
3:00 272.1 F 29.3
4:00 272.1 F 29.8

Min 271.5 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 219.2 F 0.0
10:00 229.6 F 0.0
15:00 227.1 F 0.0
20:00 224.7 F 0.0
25:00 224.4 F 0.0
30:00 223.5 F 0.0

Drying Complete

FILLING: 2:31
HEATING: 18:42
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:30
CYCLE COMPLETE

05/15/2023 04:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____