
Midmark M11 - v1.0.0
Total Cycles: 1031

Name: RITTER 1
Sterilizer ID: V2178198
07/20/2021 08:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.4 F 0.0
2:00 138.2 F 0.5
4:00 148.8 F 2.1
6:00 175.1 F 5.1
8:00 195.9 F 9.5
10:00 217.1 F 8.2
12:00 226.6 F 8.0
14:00 238.0 F 12.7
16:00 235.8 F 9.4
18:00 240.0 F 9.7
20:00 245.8 F 13.2
22:00 257.3 F 19.4
24:00 267.9 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.6 F 29.1
2:00 271.6 F 28.3
3:00 272.3 F 29.1
4:00 272.3 F 28.8

Min 271.3 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 211.3 F 0.0
10:00 201.5 F 0.0
15:00 200.5 F 0.0
20:00 198.7 F 0.0
25:00 218.0 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:23
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:03:29
CYCLE COMPLETE

07/20/2021 09:26 AM
ACCEPTED: X
REJECTED:
Operator: U01