
Midmark M11 - v1.0.0
Total Cycles: 4140

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2023 01:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.9 F 0.0
2:00 163.0 F 3.3
4:00 178.6 F 7.4
6:00 203.1 F 12.2
8:00 222.2 F 6.0
10:00 231.3 F 10.9
12:00 231.8 F 7.4
14:00 245.1 F 13.4
16:00 243.7 F 11.9
18:00 256.1 F 18.4
20:00 267.6 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.5 F 28.6
2:00 271.3 F 28.0
3:00 272.5 F 28.7
4:00 271.6 F 28.0

Min 271.2 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.8 F 0.0
5:00 238.5 F 0.0
10:00 219.9 F 0.0
15:00 227.9 F 0.0
20:00 219.6 F 0.0
25:00 218.4 F 0.0
30:00 210.9 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 21:23
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:00:00
CYCLE COMPLETE

12/20/2023 02:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____