
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007416
7:42 24/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:00
STERILIZING 2:59
136.4 C 222kPa 2:59
136.4 C 224kPa 3:29
136.4 C 220kPa 3:59
136.4 C 223kPa 4:29
136.4 C 221kPa 4:59
136.4 C 222kPa 5:29
136.4 C 217kPa 5:59
136.4 C 224kPa 6:29
136.4 C 224kPa 6:29
VENTING 6:30
AIR DRYING 7:18
CYCLE COMPLETE 37:18

Digital Signature #
B8F5FC3FE1FC55B4
