

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1716

Name: RITTER2  
Sterilizer ID: V2313228  
11/14/2022 12:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.4 F 0.0  
2:00 186.7 F 1.7  
4:00 199.7 F 6.5  
6:00 218.1 F 12.0  
8:00 226.8 F 8.1  
10:00 236.7 F 11.7  
12:00 241.1 F 12.1  
14:00 241.0 F 11.2  
16:00 254.8 F 18.2  
18:00 266.5 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.5 F 29.5  
2:00 271.4 F 28.7  
3:00 272.3 F 29.7  
4:00 272.1 F 29.1  
  
Min 271.3 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.4 F 0.0  
5:00 218.0 F 0.0  
10:00 240.9 F 0.0  
15:00 233.7 F 0.0  
20:00 223.4 F 0.0  
25:00 217.9 F 0.0  
30:00 214.2 F 0.0  
Drying Complete

FILLING: 1:31  
HEATING: 19:29  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:08  
CYCLE COMPLETE

11/14/2022 01:05 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_