
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006372
14:42 27/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:18
STERILIZING 4:25
136.5 C 224kPa 4:25
135.9 C 222kPa 4:55
136.2 C 222kPa 5:25
136.0 C 222kPa 5:55
136.3 C 224kPa 6:25
136.4 C 223kPa 6:55
136.3 C 220kPa 7:25
136.1 C 221kPa 7:55
136.1 C 221kPa 7:55
VENTING 7:55
AIR DRYING 8:42
CYCLE COMPLETE 10:01

Digital Signature #
F8E166DE9FB54D13
