
Midmark M11 - v1.0.0
Total Cycles: 643

Name: MIDMARK
Sterilizer ID: V2178198
02/01/2021 07:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 74.6 F 0.0
4:00 93.9 F 0.1
6:00 104.7 F 0.2
8:00 135.3 F 0.3
10:00 174.6 F 1.7
12:00 210.2 F 5.0
14:00 219.3 F 8.5
16:00 222.3 F 12.4
18:00 226.8 F 7.0
20:00 237.1 F 11.7
22:00 235.6 F 8.5
24:00 240.7 F 10.4
26:00 246.4 F 13.1
28:00 259.0 F 19.9
30:00 269.8 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 27.9
1:00 272.4 F 28.4
2:00 271.5 F 28.1
3:00 272.0 F 28.2
4:00 271.8 F 28.4

Min 271.1 F 27.7
Max 272.6 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.7 F 0.0
5:00 230.3 F 0.0
10:00 229.7 F 0.0
15:00 226.1 F 0.0
20:00 205.5 F 0.0
25:00 195.2 F 0.0
30:00 187.5 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:57
STERILIZING: 4:00
VENTING: 2:54
DRYING: 30:00
TOTAL CYCLE: 01:08:39
CYCLE COMPLETE

02/01/2021 08:39 AM
ACCEPTED: X
REJECTED:
Operator: U01