
Midmark M11 - v1.0.5
Total Cycles: 1098

Name: RITTER2
Sterilizer ID: V2313228
05/11/2022 03:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.9 F 0.0
2:00 192.9 F 4.0
4:00 205.5 F 9.5
6:00 223.2 F 7.1
8:00 235.1 F 11.1
10:00 235.3 F 9.2
12:00 238.8 F 9.7
14:00 249.3 F 15.1
16:00 261.4 F 22.0
18:00 272.4 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.4 F 29.1
2:00 271.8 F 29.1
3:00 272.4 F 29.0
4:00 271.7 F 29.0

Min 271.5 F 28.6
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.6 F 0.0
5:00 212.4 F 0.0
10:00 210.3 F 0.0
15:00 207.7 F 0.0
Dry Cycle Aborted

FILLING: 2:38
HEATING: 18:26
STERILIZING: 4:00
VENTING: 3:11
DRYING: 16:43
TOTAL CYCLE: 00:44:58

05/11/2022 04:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____