
Midmark M11 - v1.0.0
Total Cycles: 4064

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2023 11:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.2 F 0.0
2:00 174.4 F 2.8
4:00 179.5 F 6.7
6:00 210.8 F 11.2
8:00 218.3 F 5.6
10:00 231.2 F 10.5
12:00 231.3 F 7.6
14:00 241.5 F 12.9
16:00 241.4 F 11.4
18:00 253.6 F 17.5
20:00 264.3 F 23.8
22:00 271.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.2 F 29.3
2:00 272.7 F 29.4
3:00 272.1 F 29.1
4:00 271.6 F 28.8

Min 271.1 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 229.7 F 0.0
10:00 237.1 F 0.0
15:00 228.1 F 0.0
20:00 228.2 F 0.0
25:00 226.0 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 22:10
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:01:25
CYCLE COMPLETE

11/29/2023 12:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____