
Midmark M11 - v1.0.0
Total Cycles: 2684

Name: RITTER 1
Sterilizer ID: V2178198
12/15/2022 01:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.2 F 0.0
2:00 166.9 F 1.2
4:00 196.5 F 4.8
6:00 208.8 F 10.2
8:00 222.9 F 5.3
10:00 232.7 F 10.1
12:00 237.2 F 8.1
14:00 244.0 F 12.9
16:00 242.5 F 11.3
18:00 255.9 F 18.2
20:00 267.5 F 25.3
22:00 272.6 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.7
1:00 272.4 F 28.9
2:00 272.8 F 29.3
3:00 271.8 F 28.7
4:00 272.7 F 29.0

Min 271.8 F 28.6
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 207.5 F 0.0
10:00 227.3 F 0.0
15:00 236.7 F 0.0
20:00 231.0 F 0.0
25:00 226.5 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 22:07
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:19
CYCLE COMPLETE

12/15/2022 02:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____