

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1908

Name: RITTER2  
Sterilizer ID: V2313228  
01/09/2023 03:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.0 F 0.0  
2:00 178.8 F 3.8  
4:00 191.2 F 8.7  
6:00 216.1 F 7.7  
8:00 225.6 F 9.5  
10:00 232.7 F 8.7  
12:00 242.2 F 12.3  
14:00 245.3 F 13.6  
16:00 258.7 F 20.9  
18:00 270.5 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.4  
1:00 272.6 F 30.0  
2:00 271.6 F 29.2  
3:00 272.4 F 30.2  
4:00 272.5 F 29.9  
  
Min 271.3 F 29.2  
Max 272.8 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.0 F 0.0  
5:00 240.1 F 0.0  
10:00 233.5 F 0.0  
15:00 226.5 F 0.0  
20:00 226.2 F 0.0  
25:00 230.0 F 0.0  
30:00 232.4 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 18:47  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:57:29  
CYCLE COMPLETE

01/09/2023 04:24 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_