
Midmark M11 - v1.0.0
Total Cycles: 5418

Name: RITTER 1
Sterilizer ID: V2178198
03/17/2025 08:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.4 F 0.0
2:00 71.1 F 0.3
4:00 94.7 F 1.7
6:00 124.2 F 4.2
8:00 147.1 F 7.2
10:00 171.6 F 10.7
12:00 187.2 F 4.0
14:00 204.2 F 4.3
16:00 215.4 F 7.4
18:00 226.4 F 11.0
20:00 227.1 F 6.5
22:00 237.8 F 10.7
24:00 234.7 F 7.8
26:00 247.0 F 13.8
28:00 258.0 F 19.8
30:00 268.1 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.7 F 29.0
2:00 271.7 F 28.4
3:00 272.8 F 29.1
4:00 271.9 F 28.4

Min 271.6 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.8 F 0.0
5:00 177.6 F 0.0
10:00 189.1 F 0.0
15:00 192.8 F 0.0
20:00 195.7 F 0.0
25:00 197.0 F 0.0
30:00 197.4 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 31:22
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:09:40
CYCLE COMPLETE

03/17/2025 09:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____