
Midmark M11 - v1.0.5
Total Cycles: 3145

Name: RITTER2
Sterilizer ID: V2313228
11/20/2023 01:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.8 F 0.0
2:00 162.0 F 2.5
4:00 183.5 F 6.4
6:00 203.4 F 11.1
8:00 215.4 F 4.4
10:00 231.1 F 9.2
12:00 236.9 F 10.4
14:00 242.5 F 12.1
16:00 241.6 F 11.2
18:00 254.8 F 18.0
20:00 266.4 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.6 F 29.1
2:00 271.3 F 28.5
3:00 272.7 F 29.2
4:00 271.7 F 28.5

Min 271.2 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.4 F 0.0
5:00 229.0 F 0.0
10:00 207.1 F 0.0
15:00 207.1 F 0.0
20:00 205.3 F 0.0
25:00 201.4 F 0.0
30:00 199.0 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 21:32
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:00:06
CYCLE COMPLETE

11/20/2023 02:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____