
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001567
9:39 27/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:09
STERILIZING 3:58
136.0 C 220kPa 3:58
136.0 C 221kPa 4:28
136.1 C 222kPa 4:58
136.4 C 225kPa 5:28
136.4 C 222kPa 5:58
136.1 C 219kPa 6:28
136.4 C 223kPa 6:58
136.1 C 219kPa 7:28
136.1 C 219kPa 7:28
VENTING 7:28
CYCLE ABORTED 8:08
STOP BUTTON PRESSED
NOT STERILE
END TIME 8:08

Digital Signature #
C88F48D3B90D2D84
