
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000110
10:25 27/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:45
STERILIZING 3:47
136.0 C 220kPa 3:47
136.3 C 220kPa 4:17
136.4 C 225kPa 4:47
136.4 C 223kPa 5:17
136.4 C 223kPa 5:47
136.2 C 223kPa 6:17
136.4 C 222kPa 6:47
136.2 C 223kPa 7:17
136.2 C 223kPa 7:17
VENTING 7:17
CYCLE ABORTED 8:05
STOP BUTTON PRESSED
NOT STERILE
END TIME 8:05

Digital Signature #
A9E7DF4207ED1BCD
