
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000851
14:01 07/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 3:00
STERILIZING 4:09
136.4 C 233kPa 4:09
136.3 C 232kPa 4:39
136.4 C 233kPa 5:09
136.2 C 232kPa 5:39
136.4 C 232kPa 6:09
136.4 C 233kPa 6:39
136.4 C 233kPa 7:09
136.4 C 232kPa 7:39
136.4 C 232kPa 7:39
VENTING 7:39
AIR DRYING 8:34
CYCLE COMPLETE 23:34

Digital Signature #
5F1707BA39714405
