

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2359

Name: RITTER 1  
Sterilizer ID: V2178198  
09/22/2022 10:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.5 F 0.0  
2:00 185.0 F 3.6  
4:00 196.7 F 8.1  
6:00 216.7 F 11.2  
8:00 229.7 F 8.5  
10:00 240.2 F 12.1  
12:00 240.1 F 11.1  
14:00 236.9 F 9.2  
16:00 251.1 F 15.9  
18:00 262.8 F 22.7  
20:00 272.3 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.5 F 29.0  
2:00 271.3 F 28.3  
3:00 272.6 F 29.1  
4:00 272.0 F 28.6  
  
Min 271.2 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.1 F 0.0  
5:00 224.4 F 0.0  
10:00 201.8 F 0.0  
15:00 204.1 F 0.0  
20:00 203.0 F 0.0  
25:00 206.4 F 0.0  
30:00 225.6 F 0.0  
Drying Complete

FILLING: 2:35  
HEATING: 20:16  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:00:03  
CYCLE COMPLETE

09/22/2022 11:34 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_