

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1707

Name: RITTER 1  
Sterilizer ID: V2178198  
03/10/2022 11:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.5 F 0.0  
2:00 151.7 F 1.6  
4:00 170.6 F 4.3  
6:00 196.6 F 8.4  
8:00 219.1 F 10.7  
10:00 226.6 F 8.0  
12:00 238.8 F 13.2  
14:00 239.1 F 10.5  
16:00 236.1 F 8.6  
18:00 251.0 F 15.6  
20:00 262.9 F 22.5  
22:00 272.1 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.2  
1:00 272.2 F 28.9  
2:00 271.7 F 28.2  
3:00 272.2 F 28.9  
4:00 272.0 F 28.5

Min 271.3 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.9 F 0.0  
5:00 195.5 F 0.0  
10:00 190.7 F 0.0  
15:00 182.7 F 0.0  
20:00 174.1 F 0.0  
25:00 169.1 F 0.0  
30:00 167.2 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 22:03  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:00:23  
CYCLE COMPLETE

03/10/2022 12:57 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02