
Midmark M11 - v1.0.5
Total Cycles: 1123

Name: RITTER2
Sterilizer ID: V2313228
05/26/2022 12:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.8 F 0.0
2:00 174.5 F 3.0
4:00 182.1 F 7.4
6:00 210.5 F 13.0
8:00 225.1 F 8.3
10:00 237.6 F 12.1
12:00 237.8 F 11.5
14:00 238.1 F 10.4
16:00 252.1 F 17.5
18:00 264.3 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.7
1:00 272.7 F 30.1
2:00 271.7 F 29.9
3:00 272.7 F 30.1
4:00 271.6 F 29.6

Min 271.6 F 29.5
Max 272.8 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.1 F 0.0
5:00 189.3 F 0.0
10:00 191.8 F 0.0
15:00 193.6 F 0.0
20:00 192.1 F 0.0
25:00 188.0 F 0.0
30:00 204.3 F 0.0

Drying Complete

FILLING: 2:03
HEATING: 19:52
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:05
CYCLE COMPLETE

05/26/2022 01:27 PM
ACCEPTED: X
REJECTED:
Operator: U02