
Midmark M11 - v1.0.5
Total Cycles: 2652

Name: RITTER2
Sterilizer ID: V2313228
07/17/2023 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.3 F 0.0
2:00 178.0 F 1.4
4:00 180.2 F 4.6
6:00 200.7 F 9.1
8:00 221.5 F 11.5
10:00 227.6 F 8.0
12:00 238.3 F 12.7
14:00 237.3 F 10.2
16:00 238.0 F 9.0
18:00 249.1 F 15.2
20:00 260.5 F 21.7
22:00 270.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.9
1:00 272.4 F 29.7
2:00 272.0 F 29.0
3:00 271.8 F 29.4
4:00 272.6 F 29.4

Min 271.3 F 28.8
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 214.1 F 0.0
10:00 216.3 F 0.0
15:00 217.0 F 0.0
20:00 222.9 F 0.0
25:00 219.9 F 0.0
30:00 221.7 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 22:38
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:01:45
CYCLE COMPLETE

07/17/2023 11:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____