
Midmark M11 - v1.0.5
Total Cycles: 934

Name: RITTER2
Sterilizer ID: V2313228
03/31/2022 07:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.2 F 0.0
2:00 207.3 F 1.4
4:00 211.5 F 6.9
6:00 231.7 F 13.0
8:00 232.5 F 9.7
10:00 237.8 F 8.9
12:00 240.9 F 11.2
14:00 246.9 F 14.4
16:00 259.7 F 21.5
18:00 271.0 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.5 F 29.8
2:00 271.4 F 29.2
3:00 272.7 F 30.1
4:00 272.1 F 29.5

Min 271.3 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 221.3 F 0.0
10:00 226.6 F 0.0
15:00 228.1 F 0.0
20:00 222.6 F 0.0
25:00 226.6 F 0.0
30:00 230.7 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 18:44
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:57:47
CYCLE COMPLETE

03/31/2022 08:54 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____