
Midmark M11 - v1.0.0
Total Cycles: 839

Name: RITTER 1
Sterilizer ID: V2178198
05/26/2021 02:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.7 F 0.0
2:00 149.6 F 1.6
4:00 161.9 F 4.6
6:00 196.2 F 9.0
8:00 214.9 F 8.2
10:00 223.7 F 9.3
12:00 233.7 F 9.4
14:00 238.7 F 11.9
16:00 237.9 F 10.1
18:00 251.6 F 16.7
20:00 263.4 F 23.5
22:00 271.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.6 F 29.5
2:00 271.1 F 28.9
3:00 272.3 F 29.6
4:00 271.4 F 28.9

Min 271.0 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 235.1 F 0.0
10:00 228.6 F 0.0
15:00 228.4 F 0.0
20:00 208.3 F 0.0
25:00 193.8 F 0.0
30:00 201.5 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 22:08
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:33
CYCLE COMPLETE

05/26/2021 03:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____