
Midmark M11 - v1.0.5
Total Cycles: 2051

Name: RITTER2
Sterilizer ID: V2313228
02/14/2023 05:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.9 F 0.0
2:00 204.5 F 3.7
4:00 221.3 F 9.3
6:00 225.1 F 7.5
8:00 237.0 F 11.7
10:00 238.1 F 10.2
12:00 237.9 F 9.5
14:00 252.6 F 17.1
16:00 265.1 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.8
1:00 272.1 F 29.6
2:00 272.0 F 29.0
3:00 271.6 F 29.2
4:00 272.5 F 29.5

Min 271.4 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 231.7 F 0.0
10:00 243.7 F 0.0
15:00 237.6 F 0.0
20:00 227.8 F 0.0
25:00 224.8 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 2:25
HEATING: 17:37
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:15
CYCLE COMPLETE

02/14/2023 06:19 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____