

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000486  
11:39 26/03/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:07  
PRESSURIZING 3:15  
STERILIZING 4:50  
136.5 C 233kPa 4:50  
136.4 C 233kPa 5:20  
136.4 C 233kPa 5:50  
136.4 C 233kPa 6:20  
136.3 C 231kPa 6:50  
136.2 C 229kPa 7:20  
136.1 C 231kPa 7:50  
136.2 C 231kPa 8:20  
136.2 C 231kPa 8:20  
VENTING 8:20  
AIR DRYING 9:14  
CYCLE COMPLETE 17:49

-----  
Digital Signature #  
E9CD09714D695C12  
-----