
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002381
16:03 13/12/2024
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 4:05
STERILIZING 4:51
136.0 C 222kPa 4:51
136.2 C 223kPa 5:21
136.4 C 225kPa 5:51
136.2 C 224kPa 6:21
135.9 C 222kPa 6:51
136.4 C 225kPa 7:21
136.4 C 226kPa 7:51
136.3 C 225kPa 8:21
136.4 C 225kPa 8:51
136.4 C 226kPa 9:21
136.2 C 222kPa 9:51
136.4 C 224kPa 10:21
136.4 C 223kPa 10:51
136.4 C 226kPa 11:21
136.4 C 226kPa 11:51
136.4 C 224kPa 12:21
136.2 C 221kPa 12:51
136.4 C 226kPa 13:21
136.3 C 225kPa 13:51
136.4 C 227kPa 14:21
136.4 C 226kPa 14:51
136.4 C 226kPa 14:51
VENTING 14:51
AIR DRYING 15:41
CYCLE COMPLETE 19:26

Digital Signature #
18E42732D722800B
