

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000391  
6:59 02/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:48  
PRESSURIZING 3:48  
STERILIZING 5:11  
136.0 C 220kPa 5:11  
136.4 C 222kPa 5:41  
136.4 C 223kPa 6:11  
136.3 C 220kPa 6:41  
136.4 C 223kPa 7:11  
136.1 C 222kPa 7:41  
136.3 C 218kPa 8:11  
136.4 C 222kPa 8:41  
136.4 C 222kPa 8:41  
VENTING 8:41  
AIR DRYING 9:54  
CYCLE COMPLETE 24:34

-----  
Digital Signature #  
3D528E95F10D3547  
-----