

-----  
Midmark M11 - v1.0.0  
Total Cycles: 982

Name: RITTER 1  
Sterilizer ID: V2178198  
07/08/2021 09:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.2 F 0.0  
2:00 191.0 F 3.6  
4:00 202.6 F 9.0  
6:00 226.3 F 6.3  
8:00 236.5 F 10.9  
10:00 234.0 F 8.1  
12:00 240.8 F 10.2  
14:00 247.4 F 13.7  
16:00 260.5 F 20.9  
18:00 272.4 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 271.9 F 28.1  
2:00 272.3 F 28.4  
3:00 272.5 F 28.8  
4:00 271.6 F 28.0  
  
Min 271.5 F 27.9  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.2 F 0.0  
5:00 230.8 F 0.0  
10:00 223.8 F 0.0  
15:00 226.8 F 0.0  
20:00 222.5 F 0.0  
25:00 224.5 F 0.0  
30:00 223.2 F 0.0  
Drying Complete

FILLING: 2:13  
HEATING: 18:24  
STERILIZING: 4:00  
VENTING: 2:54  
DRYING: 30:00  
TOTAL CYCLE: 00:57:31  
CYCLE COMPLETE

07/08/2021 10:02 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02