
Midmark M11 - v1.0.5
Total Cycles: 3902

Name: RITTER2
Sterilizer ID: V2313228
08/29/2024 08:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.8 F 0.0
2:00 70.0 F 0.0
4:00 75.6 F 0.0
6:00 110.4 F 0.1
8:00 148.3 F 1.0
10:00 172.6 F 3.9
12:00 196.0 F 7.0
14:00 222.0 F 11.1
16:00 219.0 F 4.7
18:00 230.1 F 8.6
20:00 241.2 F 12.9
22:00 237.7 F 9.9
24:00 238.1 F 9.2
26:00 247.7 F 14.2
28:00 259.0 F 20.4
30:00 269.4 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.2 F 28.9
2:00 272.3 F 29.3
3:00 271.5 F 28.4
4:00 272.4 F 29.1

Min 271.2 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.5 F 0.0
5:00 225.6 F 0.0
10:00 215.6 F 0.0
15:00 209.5 F 0.0
20:00 175.7 F 0.0
25:00 175.0 F 0.0
30:00 173.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:01
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:09:06
CYCLE COMPLETE

08/29/2024 09:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____