
Midmark M11 - v1.0.0
Total Cycles: 3222

Name: RITTER 1
Sterilizer ID: V2178198
05/12/2023 07:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.4 F 0.0
2:00 77.2 F 0.0
4:00 96.6 F 0.0
6:00 114.9 F 0.0
8:00 141.0 F 1.1
10:00 167.9 F 4.0
12:00 190.9 F 7.0
14:00 212.6 F 10.4
16:00 226.2 F 8.1
18:00 225.6 F 7.1
20:00 235.3 F 11.0
22:00 232.1 F 7.2
24:00 242.5 F 12.2
26:00 239.1 F 9.9
28:00 251.4 F 15.8
30:00 261.7 F 21.8
32:00 271.2 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.2 F 28.6
2:00 272.1 F 28.9
3:00 271.5 F 28.2
4:00 272.4 F 28.9

Min 271.3 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.8 F 0.0
5:00 222.7 F 0.0
10:00 216.3 F 0.0
15:00 234.0 F 0.0
20:00 228.4 F 0.0
25:00 228.6 F 0.0
30:00 228.3 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:38
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:10:41
CYCLE COMPLETE

05/12/2023 09:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____